SAP DIR: 80016490 Version: J Page 1 of 2



## **DECLARATION OF CONFORMITY**

(in accordance with ISO/IEC 17050-1)

We declare, under our sole responsibility, that the product listed below conforms to the provisions of:

- the European Council Directive 93/42/EEC of 14 June 1993 concerning medical devices
- the Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment (RoHS)

(ROID)			
Manufacturer's Name and Business Address:	Welch Allyn, Inc. 4341 State Street Road Skaneateles Falls, NY 13153, USA		
EC REP	Regulatory Affairs Representative Welch Allyn Limited Navan Business Park Dublin Road Navan, County Meath Republic of Ireland		
Product Name <sup>1,3</sup> :	Retinoscope		
<b>REF</b> 1,3	901024, RETINOSCOPE		
# 1,3	18240, 18245, 18300, 18342-VC, 18344-V		
Radio equipment <sup>2</sup> :	Not Applicable		
Object of the declaration <sup>2</sup> :	Not Applicable		
Accessories and components <sup>2</sup> :	Not Applicable		
Medical Device Conformity Assessment Route Annex <sup>1</sup> :	VII		
Medical Device Classification <sup>1</sup> :	I		
Medical Device Classification Rules <sup>1</sup> :	12		
GMDN Code and Term <sup>1</sup> :	32712 - Retinoscope, battery-powered		

<sup>&</sup>lt;sup>1</sup> applicable to the medical devices directive, 93/42/EEC

<sup>&</sup>lt;sup>2</sup> applicable to the radio equipment directive, 2014/53/EU

<sup>&</sup>lt;sup>3</sup> applicable to the RoHS directive, 2011/65/EU

## WelchAllyn<sup>\*</sup>

SAP DIR: 80016490

Version: J Page 2 of 2

UMDNS Code and Term <sup>1</sup> :	17840 - Retinoscope		
Standards Applied (Standards are applicable to the medical device directive, unless otherwise indicated):	Number	Title	
	EN 50581	Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances	
	EN 60601-1	Medical Electrical equipment – Part 1: General requirements for Safety	
	EN 60601-1-2	Medical electrical equipment – Part 1-2: General requirements for safety – Collateral Standard Safety requirements for medical electrical systems	

Authorised Signatory:

From Butter Fiona Butler, Manager Regulatory Affairs

 $\frac{2019-07-25}{\text{Date}} \quad \frac{\text{Navan}}{\text{Place of Issue}}$ 

{EU Authorised Representative}

applicable to the medical devices directive, 93/42/EEC
 applicable to the radio equipment directive, 2014/53/EU
 applicable to the RoHS directive, 2011/65/EU